

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D2047551	(X3) Date Survey Completed 01/08/2021
Name of Provider or Supplier Skincare Drs Pa	Street Address, City, State 7373 France Ave So Suite 304, Edina, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with laboratory personnel, the laboratory failed to ensure 1 of 2 Histopathology tissue marking materials was not used after the expiration date had been exceeded. Findings are as follows: 1. The laboratory performed Mohs Micrographic Surgery under the subspecialty Histopathology as confirmed by the Histology Technician (HT) during a tour of the laboratory at 1:05 p.m. on 01/08/21. 2. An expired tissue marking material was observed as present and available for use during the tour of the laboratory. See below. Cancer Diagnostics Tissue Marking Dye Lot 7137 Expiration 08/18 3. The laboratory performed approximately 257 Mohs procedures annually as indicated on the Form CMS-116 completed for the 01/08/21 survey. 4. In an interview at 1:05 p.m. on 01/08/21, the HT confirmed the above finding.</p>