

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D2059976	(X3) Date Survey Completed 12/01/2022
Name of Provider or Supplier M Health Fairview Dermatology Bloomington	Street Address, City, State 600 W 98th St E Ste 315, Bloomington, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with laboratory personnel, the laboratory failed to ensure one of two bottles of Nuclear Fast Red stain used for Histopathology tissue processing was not used after the expiration date had been exceeded in 2022. Findings are as follows: 1. The laboratory performed Mohs micrographic surgery under the subspecialty of Histopathology as confirmed by the Clinic Manager during a tour of the laboratory at 9:10 a.m. on 12/01/22. 2. The laboratory performed the Mart Immunohistochemical staining protocol for melanomas as confirmed by tissue processing staff during an interview at 9:15 a.m. on 12/01/22. 3. One of two Nuclear Fast Red stain bottles observed in the flammable materials cabinet during the tour was expired. See below. Stain Lot number Expiration Date Nuclear Fast Red 125636 2022-7-31 4. The expired stain was available for use by the laboratory August through November 2022. 5. In an interview at 9:15 a.m. on 12/01/22, tissue processing staff confirmed the above finding. .</p>