Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	24D2059976	02/27/2025
Name of Provider or Supplier	**Street Address, City, State**	
M Health Fairview Dermatology Bloomington	600 W 98th St E Ste 315, Bloomington, MN	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	. The M Health Fairview Dermatology Bloomington laboratory was found to be out of compliance with the regulations of the Clinical Laboratory Improvement Amendments of 1988 (42 C.F.R. part 493) upon completion of the recertification survey performed on February 27, 2025. The following standard-level deficiencies were cited: 493.1236 Evaluation of proficiency testing performance .
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the laboratory failed to verify the accuracy of the single Histopathology test performed in the laboratory at least twice annually in 2023 and 2024. Findings are as follows: 1. The laboratory performed Mohs micrographic surgery with microscopic examination under the specialty of Histopathology as confirmed by the Laboratory Director during a tour of the laboratory at 1:05 p.m. on February 27, 2025. 2. The laboratory was required to send three Mohs cases and one biopsy case to a peer for biannual comparison in April and November as established in the Proficiency Testing policy found in the Proficiency Peer Review section of the Lab Manual. 3. Documentation of biannual comparisons for 2023 and 2024 was not found during review of laboratory records. The laboratory was unable to provide the missing documentation. 4. The laboratory performed 780 Mohs microscopic examinations annually as indicated on the Form CMS-2567 provided by the laboratory on the date of survey. 5. In an interview at 4:10 p.m. on February 27, 2025, the Patient Care Supervisor confirmed the above findings. .