

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D2075059	(X3) Date Survey Completed 01/07/2021
Name of Provider or Supplier My Dermatologist	Street Address, City, State 5565 Blaine Avenue #200, Inver Grove Heights, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: . Based on observation and interview with laboratory personnel, the laboratory failed to ensure a stain used for viral microscopic examinations and a stain used for histopathology tissue processing were not used after the expiration date had been exceeded. Findings are as follows: 1. The laboratory performed Tzanck microscopic examinations and Mohs Histopathology slide examination as confirmed by Histotechnician 1 during review of the Laboratory Personnel Report (From CMS-209) at 10:30 a.m. on 01/07/21. 2. Instruction not to use reagents after the expiration date was established in the Tzanck (Cytodiagnostic) Smear and the Mohs Surgery procedures located in the CLIA Manual New 3-ring binder. 3. One bottle of expired Fungal Tzanck Stain was observed adjacent to the microscope and available for use during a tour of the laboratory at 10:05 a.m. on 01/07/21. See below for detailed information. Stain Exp. Lot number Tzanck 02/19/16 4050 4. One open bottle and one unopened bottle of expired Vintage Plus Hematoxylin stain were observed in the flammable liquid storage cabinet and available for use during the tour. See below for detailed information. Stain Exp. Lot number Hematoxylin 09/01/20 078985 5. In an interview at 10:05 a.m. and 10:15 a.m. on 01/07/21, the Clinic Administrator confirmed the above finding.</p>