

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  24D2093889	<b>(X3) Date Survey Completed</b>  12/20/2022
<b>Name of Provider or Supplier</b>  Kingdom Pathways	<b>Street Address, City, State</b>  426 Oxford Street, Saint Paul, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5211</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: . Based on document review and interview with laboratory personnel, the laboratory failed to investigate one unacceptable Chemistry proficiency testing (PT) results out of twelve challenges completed in 2022. Findings are as follows: 1. The laboratory performed Chemistry (Toxicology) testing as confirmed by the Technical Consultant during a tour of the laboratory at 10:05 a.m. on 12/20/22. 2. The laboratory performed PT using the American Proficiency Institute (API) provider. 3. The laboratory received one unacceptable PT result of twelve Opiate testing challenges completed in 2022 as indicated in API reports. See below. 2022 Chemistry - Miscellaneous - 1st event Test: Opiates Sample: UDS-02 Laboratory Result: Negative API expected range: Positive 4. Investigation of unacceptable PT results was required as established in the laboratory's Proficiency Testing procedure located in the Proficiency testing manual. 5. Investigation of the unacceptable PT results was not found during review of laboratory records. The laboratory was unable to provide investigation documentation upon request. 6. In an interview at 10:40 a.m. on 12/20/22, the Laboratory Director confirmed the above finding. .</p>