

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  24D2110590	<b>(X3) Date Survey Completed</b>  03/20/2018
<b>Name of Provider or Supplier</b>  Lakes Dermatology	<b>Street Address, City, State</b>  14001 Ridgedale Dr Suite 300, Minnetonka, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by:                      . Based on observation, document review and interview with laboratory personnel, the laboratory failed to ensure a histopathology staining solution was not used after the expiration date had been exceeded. Findings are as follows: 1. The laboratory performed Mohs micrographic surgery under the specialty of Histopathology as confirmed by Mohs Technician 1 (MT1) during a tour of the laboratory on 3/20/18 at 10:05 a.m. 2. One expired Stat Lab Hematoxylin staining solution jug was observed as present and available for use in the laboratory. Lot number = 48325 Expiration date = 2/1/2018 3. In an interview on 3/20/18 at 12:30 p.m., MT1 confirmed the expired Hematoxylin staining solution was available for use.</p>