

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  24D2140084	<b>(X3) Date Survey Completed</b>  01/27/2022
<b>Name of Provider or Supplier</b>  Clarus Dermatology-Stellis Health Buffalo Clinic	<b>Street Address, City, State</b>  1700 Highway 25 North, Buffalo, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D3009</b>	<p>FACILITIES CFR(s): 493.1101(c)</p> <p>The laboratory must be in compliance with applicable Federal, State, and local laboratory requirements.</p> <p>This STANDARD is not met as evidenced by:                      . Based on observation and interview with laboratory personnel, the laboratory failed to dispose excess Histopathology tissue in accordance with applicable Federal, State, and local requirements from 02/15/20 through 01/27/22. Findings are as follows: 1. The laboratory performed Histopathology testing for Mohs micrographic surgery as confirmed by the Mohs Technician (MT) during a tour of the laboratory at 1:10 p.m. on 01/27/22. 2. Instruction to dispose excess tissue as biohazardous waste according to Federal, State, and local laws was not found in the Stellis Mohs Lab Manual. 3. A biohazard container for excess tissue disposal was not observed in the laboratory during the tour. 4. The laboratory performed an estimated 350 Mohs cases annually as indicated on the CLIA application Form CMS-116 completed on date of survey, 01/27/22. 5. In an interview at 1:20 p.m. on 01/27/22, the MT confirmed the above finding.</p>