

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 24D2254382	(X3) Date Survey Completed 09/22/2022
Name of Provider or Supplier Dermatology Duluth, Llc	Street Address, City, State 4316 Rice Lake Rd Ste 101, Duluth, MN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by:</p> <p>. Based on observation and interview with laboratory personnel, the laboratory failed to establish and follow safety procedures for the disposal of excess Histopathology tissue in accordance with applicable Federal, State, and/or local requirements from 04/19/22 through 09/22/22. Findings are as follows: 1. The laboratory performed Histopathology testing for Mohs micrographic surgery as confirmed by the Mohs Technician (MT) during a tour of the laboratory at 10:00 a.m. on 09/22/22. The MT indicated testing began on 04/19/22. 2. Instruction to dispose excess tissue as biohazardous waste according to Federal, State, and/or local laws was not found in the Mohs Procedure Manual. 3. A biohazard container for excess tissue disposal was not observed in the laboratory during the tour. 4. The laboratory performed approximately 100 Mohs cases in 2022 as indicated in laboratory records on 09/22/22. 5. In an interview at 10:15 a.m. on 09/22/22, the MT confirmed excess tissue was not disposed as biohazardous waste.</p>