

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0029772	(X3) Date Survey Completed 06/09/2021
Name of Provider or Supplier Magee General Hospital	Street Address, City, State 300 3rd Ave Se, Magee, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on review of the Centers for Medicare and Medicaid Services (CMS) data system for proficiency testing, the laboratory's proficiency testing (PT) records since the last survey on 9/12/18, and confirmation by the technical consultant, the laboratory failed to enroll in an HHS approved proficiency testing program for human chorionic gonadotropin (HCG) testing, performed on the Ortho Clinical Diagnostics Vitros 5600 chemistry system, since 9/12/18. Findings include: Review of the CMS data system for proficiency testing and the laboratory's PT records since the last survey on 9/12/18 revealed no documentation of the performance of proficiency testing for HCG testing, performed on the Ortho Clinical Diagnostics Vitros 5600 chemistry system. The technical consultant listed on the CMS 209 personnel form confirmed the laboratory's annual HCG patient test volume is 122.</p>