Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 25D0030383	**(X3) Date Survey Completed** 03/10/2022
Name of Provider or Supplier Longest Student Health Center Lab	**Street Address, City, State** 360 Hardy Road, Mississippi State, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5449	CONTROL PROCEDURES CFR(s): 493.1256(d)(3)(ii)(g) Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- At least once a day patient specimens are assayed or examined perform the following for-- Each qualitative procedure, include a negative and positive control material; (g) The laboratory must document all control procedures performed. This STANDARD is not met as evidenced by: Based on review of quality control records for Neisseria gonorrhoeae (NG) and Chlamydia trachomatis (CT) testing with the Cepheid GeneXpert test system from 10/5/20 through 3/7/22, lack of documentation of establishment of an Individualized Quality Control Plan (IQCP), and interview with the technical consultant on 3/10/22 at 3:30 p.m., the laboratory failed to include a positive and negative control each day of patient testing from the date NG/CT testing began on 10/13/20 through 3/7/22. The laboratory's annual volume for NG and CT testing is 2,122. Findings include: Review of quality control records for Neisseria gonorrhoeae and Chlamydia trachomatis testing with the Cepheid GeneXpert test system from 10/5/20 through 3/7/22 revealed positive and negative controls were performed monthly on the following days: 10/5/20, 11/2/20, 12/7/20, 1/4/21, 2/1/21, 3/1/21, 4/5/21, 5/3/21, 6/7/21, 7/6/21, 8/2/21, 9/8/21, 10/4/21, 11/1/21, 12/6/21, 1/4/22, 2/7/22, 3/7/22. The technical consultant confirmed in an interview on 3/10/22 at 3:30 p.m. that positive and negative controls were not performed each day of patient testing and an IQCP had not been established for Neisseria gonorrhoeae and Chlamydia trachomatis testing with the Cepheid GeneXpert test system.