

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0316543	(X3) Date Survey Completed 08/13/2026
Name of Provider or Supplier Family Medicine Group Of Oxford	Street Address, City, State 1397 Belk Blvd, Oxford, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5221	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d) All proficiency testing evaluation and verification activities must be documented. This STANDARD is not met as evidenced by: Based on review of Proficiency Testing (PT) records and confirmation with the Testing Personnel (TP), there was no documentation of corrective action taken for unacceptable scores for four of six Chemistry Core Proficiency Testing Events reviewed. Findings include: 1. Review of PT records from 2024 through 2026 revealed the laboratory did not document corrective action taken for the following unacceptable scores: (1) Unacceptable results on Chemistry Core Event 3 in 2024 a. Co2 - 60% b. Total Iron Binding Capacity - 0% c. Direct Bilirubin - 40% (2) Unacceptable results on Chemistry Core Event 1 in 2025 a. Direct Bilirubin - 60% b. Folate - 60% (3) Unacceptable results on Chemistry Core Event 2 in 2025 a. Albumin - 60% b. Hemoglobin A1C - 60% c. Vitamin D - 50% (4) Unacceptable results on Chemistry Core Event 3 in 2025 a. CK Iso - 60 % b. Direct Bilirubin - 40% c. Vitamin D - 0% 2. An interview with testing personnel on 8/13/2026 at 12:00 p.m. confirmed the laboratory did not perform corrective action for the unacceptable scores for four of six Chemistry Core Proficiency Testing Events reviewed.