

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0317068	(X3) Date Survey Completed 09/26/2018
Name of Provider or Supplier Dr George L Cain Jr Md Pllc	Street Address, City, State 506 Alcorn Drive, Corinth, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory personnel documented annual evaluations/competency and interview with staff, the Laboratory Director had not followed written policies to assess the technical consultant competency at least annually since the laboratory's last survey, 4/12/17. Findings include: There was no documentation of a competency or evaluation for the technical consultant reflecting the duties specific for that position (see D6049 in the regulations) for the year 2017 through the day of survey (9/26/2018).</p>