

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0317885	(X3) Date Survey Completed 10/24/2023
Name of Provider or Supplier Yazoo Family Medicine Pa Dba Yazoo City Med Clinic	Street Address, City, State 805 East 15th Street, Yazoo City, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, review of laboratory personnel records since 1/4/2022, and an interview with the technical consultant (TC) at 2:30 p.m. on 10/24/2023, the laboratory failed to follow written policies to assess the technical consultant (TC) competency annually. The technical consultant competency had not been performed by the laboratory director since the TC was hired in February 2022. Findings include: 1. There was no documented competency evaluation available for review on the TC on the day of survey. 2 Laboratory policy stated that the TC would be evaluated at least annually. 3. The TC confirmed in an interview at 2:30 p.m. on 10/24/2023 that the laboratory failed to follow policies to assess technical consultant competency annually.</p>