

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0320210	(X3) Date Survey Completed 10/10/2022
Name of Provider or Supplier Monroe Regional Hospital	Street Address, City, State 400 S Chestnut St, Aberdeen, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The following condition level deficiencies were cited: D2016 - 42 C.F.R. 493.803 Condition: Successful participation, proficiency testing
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the American Proficiency Institute and the CASPER report 0153D /0155D from the Centers for Medicare and Medicaid Services data system) on 10/10 /2022, the laboratory failed to maintain satisfactory performance in two of two testing events (1st event of 2022 and 2nd event of 2022) resulting in unsuccessful performance for Blood Cell Identification. Refer to D2130.</p>

D2130

HEMATOLOGY

CFR(s): 493.851(f)

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the American Proficiency Institute and CASPER reports 0153D/0155D from the Centers for Medicare and Medicaid Services data system) on 10/10/2022, the laboratory has not successfully performed proficiency testing for Blood Cell Identification in two of two testing events. Findings include: A review of the laboratory records from the American Proficiency Institute (API) and the CMS CASPER reports 0153D/0155D revealed the laboratory scored the following for Blood Cell Identification: Year 2022 - 1st Event: 60% Year 2022 - 2nd Event: 0%