

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0651752	(X3) Date Survey Completed 01/23/2018
Name of Provider or Supplier Progressive Medical Mgmt Dba Panola Medical Center	Street Address, City, State 303 Medical Center Drive, Batesville, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on Surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the proficiency testing provider) on 1/23/2018, the laboratory had not successfully participated in proficiency testing for ABO/RHO. Findings include: Our records indicate the following proficiency testing scores for your laboratory for ABO/RHO. PROFICIENCY TESTING PROVIDER: American Proficiency Institute ABO/RHO: Year 2017 1st Event 90% Year 2017 3rd Event 90% Scores less than

	100% for this analyte or parameter indicate failure for unsatisfactory performance. A failure of the analyte or parameter for two consecutive or two out of three testing events is scored as initial unsuccessful performance.
D2163	ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(g) Failure to achieve an overall testing event score of satisfactory for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on Surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the proficiency testing provider) on 1/23/2018, the laboratory had not successfully participated in proficiency testing for ABO/RHO. Findings include: Our records indicate the following proficiency testing scores for your laboratory for ABO/RHO. PROFICIENCY TESTING PROVIDER: American Proficiency Institute ABO/RHO: Year 2017 1st Event 90% Year 2017 3rd Event 90% Scores less than 100% for this analyte or parameter indicate failure for unsatisfactory performance. A failure of the analyte or parameter for two consecutive or two out of three testing events is scored as initial unsuccessful performance.