

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D0651950	(X3) Date Survey Completed 11/19/2018
Name of Provider or Supplier Perry County General Hospital Lab	Street Address, City, State 206 Bay Avenue, Richton, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on Surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the proficiency testing provider and the Centers for Medicare and Medicaid data system) on 11/19/18, the laboratory has not successfully participated in proficiency testing for ALCOHOL (BLOOD). Findings include: Our records indicate the following proficiency testing scores for your laboratory for ALCOHOL (BLOOD). PROFICIENCY TESTING PROVIDER: American Association of Bioanalysts ALCOHOL (BLOOD): Year 2018 1st Event 0% Year 2018 2nd Event 0% Year 2018 3rd Event 0% Scores less than 80% for these analytes or parameters indicate failure for unsatisfactory performance. A failure of the analyte or parameter

	for two consecutive or two out of three testing events is scored as initial unsuccessful performance. Repeat failures are scored as non-initial unsuccessful performance.
D2118	TOXICOLOGY CFR(s): 493.845(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on Surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the proficiency testing provider and the Centers for Medicare and Medicaid data system) on 11/19/18, the laboratory has not successfully participated in proficiency testing for ALCOHOL (BLOOD). Findings include: Our records indicate the following proficiency testing scores for your laboratory for ALCOHOL (BLOOD). PROFICIENCY TESTING PROVIDER: American Association of Bioanalysts ALCOHOL (BLOOD): Year 2018 1st Event 0% Year 2018 2nd Event 0% Year 2018 3rd Event 0% Scores less than 80% for these analytes or parameters indicate failure for unsatisfactory performance. A failure of the analyte or parameter for two consecutive or two out of three testing events is scored as initial unsuccessful performance. Repeat failures are scored as non-initial unsuccessful performance.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the proficiency testing provider and the Centers for Medicare and Medicaid Services data system) on 11/19/18, the laboratory director has not provided overall management and direction in accordance with 493.1407 of this subpart.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on surveyor desk review of the laboratory proficiency testing (PT) records (graded copies from the proficiency testing provider and the Centers for Medicare and Medicaid Services data system) on 11/19/18, the laboratory director has failed to

ensure the proficiency testing samples were tested as required under subpart H of this part.