

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  25D0927282	<b>(X3) Date Survey Completed</b>  03/09/2022
<b>Name of Provider or Supplier</b>  Central Mississippi Correctional Facility	<b>Street Address, City, State</b>  3794 Highway 468, Pearl, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2007</b>	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing (PT) records for 2020, 2021 and 1st event of 2022, surveyor review of the Centers for Medicare &amp; Medicaid Services (CMS) 209 personnel form, and confirmation by the laboratory manager at 2:00 p.m. on 3/9/22, the laboratory failed to allow testing personnel (TP) who routinely perform Troponin I testing on patient samples to participate in proficiency testing events since 2019. Findings include: 1. Review of proficiency records since the last survey on 10/2 /19 revealed all 3 events of 2020, 2021 and the 1st event of 2022 were performed by laboratory personnel who do not perform Troponin I testing on patient specimens. TP #1, #2, and #3 as listed on the CMS 209 form perform all Troponin I testing on patients but did not participate in proficiency testing events. 2. According to an interview with the laboratory manager at 2:00 p.m. on the day of survey, the laboratory personnel that performed all proficiency testing events in 2020, 2021 and 2022 only perform waived testing. All Troponin I testing on patient samples is performed by the TP listed on the CMS-209. None of the TP listed on the CMS-209 participated in the proficiency testing for Troponin I.</p>
<b>D6000</b>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p>

This CONDITION is not met as evidenced by:

Based on the review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, lack of qualifying documentation, and interview by phone with the Correctional Facility Medical Director at 1:30 p.m. on 3/9/22, the laboratory does not have a laboratory director who meets the qualification requirements of 493.1405 of this subpart.

**D6003**

**LABORATORY DIRECTOR QUALIFICATIONS**

CFR(s): 493.1405 AND 493.1406

The laboratory director must be qualified to manage and direct the laboratory personnel and the performance of moderate complexity tests and must be eligible to be an operator of a laboratory within the requirements of subpart R of this part. (a) The laboratory director must possess a current license as a laboratory director issued by the State in which the laboratory is located, if such licensing is required; and (b) The laboratory director must-- (b)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(1)(ii) Be certified in anatomic or clinical pathology, or both, by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (b)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the Laboratory is located; and (b)(2)(ii) Have had laboratory training or experience consisting of: (b)(2)(ii)(A) At least one year directing or supervising non-waived laboratory testing; or (b)(2)(ii)(B) Beginning September 1, 1993, have at least 20 continuing medical education credit hours in laboratory practice commensurate with the director responsibilities defined in 493.1407; or (b)(2)(ii)(C) Laboratory training equivalent to paragraph (b)(2)(ii)(B) of this section obtained during medical residency. (For example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine); or (b)(3) Hold an earned doctoral degree in a chemical, physical, biological, or clinical laboratory science from an accredited institution; and (b)(3)(i) Be certified by the American Board of Medical Microbiology, the American Board of Clinical Chemistry, the American Board of Bioanalysis, or the American Board of Medical Laboratory Immunology; or (b)(3)(ii) Have had at least one year experience directing or supervising non-waived laboratory testing; (b)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; (b)(4)(ii) Have at least one year of laboratory training or experience, or both in non-waived testing; and (b)(4)(iii) In addition, have at least one year of supervisory laboratory experience in non-waived testing; or (b)(5)(i) Have earned a bachelor's degree in a chemical, physical, or biological science or medical technology from an accredited institution; (b)(5)(ii) Have at least 2 years of laboratory training or experience, or both in non-waived testing; and (b)(5)(iii) In addition, have at least 2 years of supervisory laboratory experience in non-waived testing; (b)(6) Be serving as a laboratory director and must have previously qualified or could have qualified as a laboratory director under 493.1406; or (b)(7) On or before February 28, 1992, qualified under State law to direct a laboratory in the State in which the laboratory is located. Laboratory director qualifications on or before February 28, 1992 The laboratory director must be qualified to manage and direct the laboratory personnel and test performance. (a) The laboratory director must possess a current license as a laboratory director issued by the State, if such licensing exists; and (b)

The laboratory director must: (b)(1) Be a physician certified in anatomical or clinical pathology (or both) by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; (b)(2) Be a physician who: (b)(2)(i) Is certified by the American Board of Pathology or the American Osteopathic Board of Pathology in at least one of the laboratory specialties; or (b)(2)(ii) Is certified by the American Board of Medical Microbiology, the American Board of Clinical Chemistry, the American Board of Bioanalysis, or other national accrediting board in one of the laboratory specialties; or (b)(2)(iii) Is certified by the American Society of Cytology to practice cytopathology or possesses qualifications that are equivalent to those required for such certification; or (b)(2)(iv) Subsequent to graduation, has had 4 or more years of full-time general laboratory training and experience of which at least 2 years were spent acquiring proficiency in one of the laboratory specialties; (b)(3) For the subspecialty of oral pathology only, be certified by the American Board of Oral Pathology, American Board of Pathology or the American Osteopathic Board of Pathology or possesses qualifications that are equivalent to those required for certification; (b)(4) Hold an earned doctoral degree from an accredited institution with a chemical, physical, or biological science as a major subject and (b)(4)(i) Is certified by the American Board of Medical Microbiology, the American Board of Clinical Chemistry, the American Board of Bioanalysis, or other national accrediting board acceptable to HHS in one of the laboratory specialties; or (b)(4)(ii) Subsequent to graduation, has had 4 or more years of full-time general laboratory training and experience of which at least 2 years were spent acquiring proficiency in one of the laboratory specialties; (b)(5) With respect to individuals first qualifying before July 1, 1971, have been responsible for the direction of a laboratory for 12 months between July 1, 1961, and January 1, 1968, and, in addition, either: (b)(5)(i) Was a physician and subsequent to graduation had at least 4 years of pertinent full-time laboratory experience; (b)(5)(ii) Held a master's degree from an accredited institution with a chemical, physical, or biological science as a major subject and subsequent to graduation had at least 4 years of pertinent full-time laboratory experience; (b)(5)(iii) Held a bachelor's degree from an accredited institution with a chemical, physical, or biological science as a major subject and subsequent to graduation had at least 6 years of pertinent full-time laboratory experience; or (b)(5)(iv) Achieved a satisfactory grade through an examination conducted by or under the sponsorship of the U.S. Public Health Service on or before July 1, 1970; or (b)(6) Qualify under State law to direct the laboratory in the State in which the laboratory is located. Note: The January 1, 1968 date for meeting the 12 months' laboratory direction requirement in paragraph (b)(5) of this section may be extended 1 year for each year of full-time laboratory experience obtained before January 1, 1958 required by State law for a laboratory director license. An exception to the July 1, 1971 qualifying date in paragraph (b)(5) of this section was made provided that the individual requested qualification approval by October 21, 1975 and had been employed in a laboratory for at least 3 years of the 5 years preceding the date of submission of his qualifications.

This STANDARD is not met as evidenced by:

Based on review of the CMS 209 personnel form, lack of qualifying documentation available for review, interview with the Correctional Facility Medical Director by phone and the laboratory manager at 1:30 p.m. on 3/9/22, the laboratory did not have an individual designated as laboratory director who meets the qualification requirements of 493.1405 of this subpart from 1/7/22 through 3/9/22.

D6033

TECHNICAL CONSULTANT-MODERATE COMPLEXITY

CFR(s): 493.1409

The laboratory must have a technical consultant who meets the qualification requirements of 493.1411 of this subpart and provides technical oversight in accordance with 493.1413 of this subpart.

This CONDITION is not met as evidenced by:

Based on the review of the CMS 209 personnel form, lack of qualifying documentation, interview by phone with the Correctional Facility Medical Director, and interview with the laboratory manager at 1:30 p.m. on 3/9/22, the laboratory does not have a technical consultant who meets the qualification requirements of 493.1411 of this subpart.

**D6035**

**TECHNICAL CONSULTANT QUALIFICATIONS**

CFR(s): 493.1411

(a) The technical consultant must be qualified and must possess a current license issued by the State in which the laboratory is located, if such licensing is required. (b) The technical consultant must-- (b)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(1)(ii) Be certified in anatomic or clinical pathology, or both, by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (b)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; and (b)(2)(ii) Have at least one year of laboratory training or experience, or both in non-waived testing, in the designated specialty or subspecialty areas of service for which the technical consultant is responsible (for example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine are qualified to serve as the technical consultant in hematology); or (b)(3)(i) Hold an earned doctoral or master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; and (b)(3)(ii) Have at least one year of laboratory training or experience, or both in non-waived testing, in the designated specialty or subspecialty areas of service for which the technical consultant is responsible; or (b)(4)(i) Have earned a bachelor's degree in a chemical, physical or biological science or medical technology from an accredited institution; and (b)(4)(ii) Have at least 2 years of laboratory training or experience, or both in non-waived testing, in the designated specialty or subspecialty areas of service for which the technical consultant is responsible. Note: The technical consultant requirements for "laboratory training or experience, or both" in each specialty or subspecialty may be acquired concurrently in more than one of the specialties or subspecialties of service, excluding waived tests. For example, an individual who has a bachelor's degree in biology and additionally has documentation of 2 years of work experience performing tests of moderate complexity in all specialties and subspecialties of service, would be qualified as a technical consultant in a laboratory performing moderate complexity testing in all specialties and subspecialties of service.

This STANDARD is not met as evidenced by:

Based on review of the CMS 209 personnel form, lack of qualifying documentation available for review, interview with the Correctional Facility Medical Director by

phone, and interview with the laboratory manager at 1:30 p.m. on 3/9/22, the laboratory did not have an individual designated as technical consultant who meets the qualification requirements of 493.1411 of this subpart from 1/7/22 through 3/9/22.

**D6049**

**TECHNICAL CONSULTANT RESPONSIBILITIES**

CFR(s): 493.1413(b)(8)(iii)

The procedures for evaluation of the competency of the staff must include, but are not limited to review of intermediate test results or worksheets, quality control records, proficiency testing results, and preventive maintenance records.

This STANDARD is not met as evidenced by:

Based on review of laboratory testing records from 10/3/19 through 3/9/22 and interview with the laboratory manager at 12:30 p.m. on 3/9/22, the following records had not been documented as reviewed by the technical consultant (TC): Findings Include: 1. Review of the laboratory records from 10/3/19 through 3/9/22 revealed the following records were not documented as reviewed by the TC: a. Temperature logs (room and refrigerator) from 1/1/21 through 3/9/22 b. Life Sign Troponin I quality control (QC) from 12/22/20 through 2/12/22 c. Proficiency Testing results for the 1st, 2nd and 3rd events of 2021 and 1st event of 2022 2. Interview with the laboratory manager at 12:20 p.m. on 3/9/22 confirmed there was no documented review of these records by the TC.