

| | | |
|----------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------|-----------------------------------------------------|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 25D1002756 | (X3) Date Survey Completed 05/27/2026 |
| Name of Provider or Supplier Oxford Pediatric Group Pllc | Street Address, City, State 101 Farm View Drive, Oxford, MS | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| D6053 | <p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory testing personnel (TP) as listed on the Centers for Medicare and Medicaid Services (CMS) 209 personnel form and interview with the Laboratory Director/Technical Consultant (LD/TC) and Laboratory Supervisor, the Technical Consultant (TC) failed to evaluate and document the performance of two of eight testing personnel at least semiannually during the first year of moderate complexity testing. Findings include: 1. Review of the laboratory personnel records since the last survey on 7/18/2024, revealed no semiannual competency evaluations were available for the performance of two of eight testing personnel. a. TP #7 initial training-08/05/2025, semiannual evaluation due-02/2026 b. TP #8 initial training-03/18/2024, semiannual evaluation due-09/2024 2. The LD/TC and Laboratory Supervisor confirmed in an interview on 5/27/2026 at 4:00 p.m., there was no semiannual competency evaluations performed on TP #7 and TP #8 by the TC.</p> |
| D6054 | <p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Thereafter, evaluations must be performed at least annually</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory testing personnel (TP) records including the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, and interview with the</p> |

Laboratory Director/Technical Consultant LD/TC and the Laboratory Supervisor, the Technical Consultant failed to evaluate the performance of six of eight testing personnel as listed on the CMS 209 form, at least annually. Findings include: 1. Review of the laboratory personnel records from 7/18/2024 through 5/27/2026 revealed no annual evaluations for the performance of six of eight testing personnel were performed by the TC. a. TP #2- competency/evaluation for 2025 and 2026 (due April 2026) b. TP #3- competency/evaluation for 2025 and 2026 (due April 2026) c. TP #4- competency/evaluation for 2025 and 2026 (due April 2026) d. TP #5- competency/evaluation for 2025 and 2026 (due May 2026) e. TP #6- competency /evaluation for 2025 and 2026 (due April 2026) f. TP #8- competency/evaluation for 2025 and 2026 (due March 2026) 3. The LD/TC and Laboratory Supervisor confirmed in an interview on 5/27/2026 at 4:00 p.m. that no annual competency evaluation had been documented as performed by the TC for six of eight testing personnel.