

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D2060779	(X3) Date Survey Completed 09/12/2019
Name of Provider or Supplier Mrh Medical Group-East Main Family Medical	Street Address, City, State 1424 E Main St, Tupelo, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5437	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(a)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must perform and document calibration procedures-- (1) Following the manufacturer's test system instructions, using calibration materials provided or specified, and with at least the frequency recommended by the manufacturer; (2) Using the criteria verified or established by the laboratory as specified in 493.1253(b) (3)-- (2)(i) Using calibration materials appropriate for the test system and, if possible, traceable to a reference method or reference material of known value; and (2)(ii) Including the number, type, and concentration of calibration materials, as well as acceptable limits for and the frequency of calibration; and (3) Whenever calibration verification fails to meet the laboratory's acceptable limits for calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Cell Dyn Emerald hematology calibration records from 8/6/17 through 9/12/19 and interview with the technical consultant at 12:30 pm on the day of survey 9/12/19, the laboratory failed to perform calibration verification on the Cell Dyn Emerald for CBC (complete blood count) parameters every 6 months as required by the manufacturer. Findings include: 1. Review of the Cell Dyn Emerald Operator's manual reveals that the manufacturer requires calibration verification on the CBC parameters every 6 months or when indicated by quality control or after major maintenance. 2. Review of the Cell Dyn Emerald calibration records revealed calibration was performed on 3/5/18 and then again on 8/14/19 which exceeds the 6 month mandatory calibration requirement. 3. Interview with the technical consultant at 12:30 pm on 9/12/19 confirmed that CBC calibration verification on the Cell Dyn Emerald was not performed every 6 months as required by the manufacturer.</p>