

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D2060779	(X3) Date Survey Completed 06/18/2025
Name of Provider or Supplier Mrh Medical Group-East Main Family Medical	Street Address, City, State 1424 E Main St, Tupelo, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2007	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>(b)(1) The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing (PT) records, review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form and interview with the Laboratory Director (LD) and Testing Personnel (TP) #1, the laboratory failed to allow testing personnel who routinely perform complete blood count (CBC) testing on patient samples to participate in five of six proficiency testing events. Findings include: 1. Review of proficiency records since the last survey revealed the 3rd event of 2023, all 3 events of 2024 and 1st event 2025 were all performed by testing TP #1. There are 3 testing personnel (TP #1, TP #2, TP #3) listed on the CMS 209 form that perform CBC testing on patients. TP #1 participated in five of six proficiency testing events. 2. An interview with the LD and TP #1, on 6/18/2025 at 3:30 p.m., revealed TP #1 performed proficiency testing for the 3rd event of 2023, all 3 events of 2024 and 1st event of 2025. The other two testing personnel did not perform proficiency testing for those events.</p>