

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D2064164	(X3) Date Survey Completed 08/04/2021
Name of Provider or Supplier Priority Medical Clinic, Llc	Street Address, City, State 601 North 15th Avenue, Laurel, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6053	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form and personnel records since the last survey on 4-3-19 and interview with the technical consultant on 8-4-21 at 11:00 a.m., the technical consultant failed to evaluate and document the performance of Testing Personnel #2 and #3, responsible for moderate complexity testing, at least semiannually during the first year these individuals tested patient complete blood count (CBC) specimens. Findings include: Review of the CMS 209 personnel form and personnel records since the last survey on 4-3-19 revealed no semiannual evaluations by the technical consultant for the performance of moderate complexity CBC testing for Testing Personnel #2, since this individual's initial evaluation on 10-6-20, and for Testing Personnel #3, since this individual's initial evaluation on 8-5-20. In an interview with the technical consultant on 8-4-21 at 11:00 a.m., the technical consultant confirmed semiannual evaluations were not performed for Testing Personnel #2 and #3.</p>