

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 25D2239259	(X3) Date Survey Completed 04/29/2026
Name of Provider or Supplier Quitman Community Hospital	Street Address, City, State 340 Getwell St, Marks, MS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6015	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)</p> <p>(e)(4) Ensure that the laboratory is enrolled in an HHS approved proficiency testing program for the testing performed and that--</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory proficiency testing records, Centers for Medicare and Medicaid Services (CMS) database proficiency testing CASPER report and interview with the laboratory manager, the laboratory director failed to ensure the laboratory enrolled and participated in an HHS approved proficiency testing (PT) program for Chemistry, Toxicology and Blood Gas testing for one of six PT events. Findings include: 1. Review of laboratory proficiency testing records from 6/27/2024 through 4/29/2026 revealed no evidence of enrollment or participation in proficiency testing for the first PT event of 2025 for Chemistry, Toxicology, and Blood Gas testing. 2. Review of the CMS database proficiency testing CASPER report revealed no scores or participation for Chemistry, Toxicology or Blood Gas testing for the 1st PT event of 2025. The laboratory failed to enroll and participate in one of six proficiency events. 3. During an interview on 4/29/2026 at 3:00 p.m., the laboratory manager confirmed the laboratory was not enrolled in proficiency testing for Chemistry, Toxicology, and Blood Gas testing in time to participate in the first PT event of 2025. THIS IS A REPEAT DEFICIENCY</p>
D6041	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(3)</p> <p>(b)(3) Enrollment and participation in an HHS approved proficiency testing program commensurate with the services offered;</p>

This STANDARD is not met as evidenced by:

Based on review of the laboratory proficiency testing records, Centers for Medicare and Medicaid Services (CMS) database proficiency testing CASPER report and interview with the laboratory manager, the technical consultant failed to ensure the laboratory enrolled and participated in an HHS approved proficiency testing (PT) program for Chemistry, Toxicology and Blood Gas testing for one of six PT events. Findings include: 1. Review of laboratory proficiency testing records from 6/27/2024 through 4/29/2026 revealed no evidence of enrollment or participation in proficiency testing for the first PT event of 2025 for Chemistry, Toxicology, and Blood Gas testing. 2. Review of the CMS database proficiency testing CASPER report revealed no scores or participation for Chemistry, Toxicology or Blood Gas testing for the 1st PT event of 2025. The laboratory failed to enroll and participate in one of six proficiency events. 3. During an interview on 4/29/2026 at 3:00 p.m., the laboratory manager confirmed the laboratory was not enrolled in proficiency testing for Chemistry, Toxicology, and Blood Gas testing in time to participate in the first PT event of 2025. THIS IS A REPEAT DEFICIENCY