

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D0439378	(X3) Date Survey Completed 06/02/2026
Name of Provider or Supplier St Louis Cancer Care	Street Address, City, State 10004 Kennerly Rd Ste 137a, Saint Louis, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on review of Sysmex XN-L Operator's Manual, laboratory humidity records for January 2026 to date June 2, 2026, and patient hematology results and interview with the testing personnel (TP) #1, the laboratory failed to ensure the humidity was maintained for 32 of 107 days. Findings: 1. Review of the Sysmex XN-L Operator's Manual states "Operating environment relative humidity: 20-85%". 2. Review of the laboratory humidity records shows that the humidity in the laboratory where the Sysmex XN-430 is located was out of range for 32 of 107 days from January 2026 to date June 2, 2026. 3. Review of patient hematology results showed the laboratory performs approximately 3,735 CBC's annually. 4. Interview with the TP #1 on June 2, 2026 at 3:30 PM confirmed the laboratory failed to ensure the humidity was maintained for 32 days.</p>