

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D0445040	(X3) Date Survey Completed 05/06/2024
Name of Provider or Supplier Advanced Dermatology & Skin Cancer Center	Street Address, City, State 1419 Village Drive, Saint Joseph, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation of tissue marking and margin dyes and interview with the practice manager, the laboratory failed to discard expired tissue marking dyes. Findings: 1. Observation of tissue marking and margin dyes showed: Violet dye lot #21270, expiration date September 30, 2023 still in use Blue dye lot #21266, expiration date October 31, 2023 still in use Three bottles of dye, with no label to include the product name, color, lot number, expiration date, open date, stil in use. 2. Interview with practice manager on May 6, 2024 at 10:00 AM confirmed the laboratory failed to discard expired tissue marking dyes.</p>