

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D0445758	(X3) Date Survey Completed 09/22/2020
Name of Provider or Supplier Freeman Neosho Hospital	Street Address, City, State 113 W Hickory St, Neosho, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation of Chemistry Nearl water and interview with technical supervisor (TS) #3, the laboratory failed to monitor expiration dates. Findings: 1. Observation of chemistry Nearl water Lot #52438 showed expiration date of 06/30 /2020 and still in use. 2. Interview with TS #3 on September 22, 2020 at 8:30 A.M. confirmed the laboratory failed to monitor expiration dates of Nearl water.</p>