

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D0447886	(X3) Date Survey Completed 01/31/2018
Name of Provider or Supplier Mercy Clinic Urology	Street Address, City, State 1965 S Fremont Suite 370, Springfield, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6063	<p>LABORATORY TESTING PERSONNEL CFR(s): 493.1421</p> <p>The laboratory must have a sufficient number of individuals who meet the qualification requirements of 493.1423, to perform the functions specified in 493.1425 for the volume and complexity of tests performed.</p> <p>This CONDITION is not met as evidenced by: Based on review of personnel records and interview with the laboratory director, the laboratory failed to have academic credentials required to qualify one of ten testing personnel performing moderate complexity testing (refer to tag #6065).</p>
D6065	<p>TESTING PERSONNEL QUALIFICATIONS CFR(s): 493.1423(b)(1)(2)(3)(4)(i)</p> <p>(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located or have earned a doctoral, master's, or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; or (b)(2) Have earned an associate degree in a chemical, physical or biological science or medical laboratory technology from an accredited institution; or (b)(3) Be a high school graduate or equivalent and have successfully completed an official military medical laboratory procedures course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(4)(i) Have earned a high school diploma or equivalent; and</p> <p>This STANDARD is not met as evidenced by: Based on review of personnel records and interview with the laboratory director, the</p>

laboratory failed to have documentation of academic credentials to qualify one of ten testing personnel performing moderate complexity testing. Findings: 1. Review of personnel records revealed the laboratory did not have academic credentials to qualify testing person # 3 for moderate complexity testing. 2. Interview with the director on January 31, 2018 at 09:30 AM confirmed the laboratory did not have academic credentials required to qualify testing person #3.