Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D0672729	(X3) Date Survey Completed 01/26/2021
Name of Provider or Supplier Jefferson City Correctional Center	**Street Address, City, State** 8200 No More Victims Rd, Jefferson City, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6046	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8) (b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently. This STANDARD is not met as evidenced by: Based on review of competency/performance evaluations and interview with the laboratory manager, the technical consultant failed to evaluate and document the competency for 23 of 23 testing personnel performing moderate complexity troponin testing for 2019 and 2020. Findings: 1. Review of competency/performance evaluations revealed no documentation to show the technical consultant evaluated and documented the competency for 23 testing personnel performing moderate complexity troponin testing. 2. Interview with laboratory manager on January 26, 2021 at 1:15 PM confirmed the technical consultant failed to evaluate and document the competency of all testing personnel for 2019 and 2020.