| **Statement of Deficiencies** | **(X1) Provider/Supplier/CLIA Identification Number** | **(X3) Date Survey Completed** |
|---|---|---|
| | 26D0678066 | 09/29/2026 |
| **Name of Provider or Supplier** | **Street Address, City, State** | |
| Crmc Abg Lab | 1600 E Evergreen (Respiratory), Cameron, MO | |

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---|---|
| **D5413** | TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT<br>CFR(s): 493.1252(b)<br><br>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.<br><br>This STANDARD is not met as evidenced by:<br>Based on observation of the arterial blood gas (ABG) laboratory refrigerator, review of ABG laboratory temperature logs for January 2026 to date September 29, 2026, patient results and interview with the testing personnel (TP) #1, the laboratory failed to ensure the refrigerator temperature was maintained for 14 of 272 days. Findings: 1. Observation of ABG laboratory refrigerator on September 29, 2026 showed 1 each of Siemens RAPIDPoint 500e Measurement Cartridge M10 lot #104914473621011117 exp 12/31/2026, 1 each of Siemens RAPIDPoint 500e Measurement Cartridge M10 lot #104914473623712813 exp 01/03/2027, 1 each of Siemens RAPIDPoint 500e Measurement Cartridge M10 lot #104914473622405925 exp 12/26/2026, 7 boxes of Eutrol epoc GAS-ISE Metabolites Level 1 lot #(10)266-1-B549 exp 05/31/2027, 8 boxes of Eutrol epoc GAS-ISE Metabolites Level 2 lot #(10) 266-2-B424 exp 04/30 /2027, and 11 boxes of Eutrol epoc GAS-ISE Metabolites Level 3 lot #(10)266-3-B550 exp 05/31/2027 with temperature storage requirements of 2 to 8 degrees Celsius. 2. Review of "Temperature Log" states acceptable range "FRIDGE TEMP 2-8 deg C". 3. Review of the ABG laboratory temperature logs showed that the refrigerator temperature was out of range for 14 of 272 days from January 2026 and to date September 29, 2026. 4. Review of patient results showed the ABG laboratory |

| | |
|---|---|
| | performs approximately 1,025 ABG patient tests annually. 5. Interview with the TP #1 on September 29, 2026 at 2:30 PM confirmed the laboratory failed to ensure the refrigerator temperature was maintained for 14 days. |