

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D0960320	(X3) Date Survey Completed 05/13/2025
Name of Provider or Supplier Midwest Regional Allergy, Asthma, Arthritis	Street Address, City, State 1027 S Main Suite 202, Joplin, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>(a) A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of procedures, and interview with the technical supervisor (TS), the laboratory failed to provide a procedure for competency assessment. Findings: 1. Review of procedures revealed no competency assessment procedure. 2. Interview with TS on May 13, 2025 at 11:00 AM confirmed the laboratory failed to provide a procedure for competency assessment.</p>
D6102	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(12)</p> <p>(e)(12) Ensure that prior to testing patients specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results;</p> <p>This STANDARD is not met as evidenced by: Based on review of competencies, and interview with the technical supervisor (TS), the laboratory director failed to perform initial training on one of one new testing personnel prior to testing patient specimens. Findings: 1. Review of competencies</p>

showed no initial training for testing personnel #3. 2. Interview with TS on May 13, 2025 at 11:00 AM confirmed the laboratory director failed to perform initial training on testing personnel #3 prior to testing patient specimens.