

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D1071136	(X3) Date Survey Completed 10/22/2019
Name of Provider or Supplier Kansas City Skin & Cancer Center	Street Address, City, State 8656 N Ambassador Drive, Kansas City, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory correlation records for 2017, 2018, 2019 and interview with the laboratory director, the laboratory failed to establish a means to verify the accuracy of AFB stain and gram stain testing twice a year. Findings: 1. Review of correlation records for 2017, 2018 and to date 2019 revealed a lack of documentation to prove accuracy 2 times a year for AFB stain testing. 2. Review of correlation records for 2017, 2018 and to date 2019 revealed a lack of documentation to prove accuracy 2 times a year for gram stain testing. 3. Interview with the laboratory director on October 22, 2019 at 11:00 AM confirmed the laboratory failed to verify the accuracy of AFB stain and gram stain testing twice annually for 2017, 2018 and to date 2019.</p>