| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number<br><br>26D1098985 | (X3) Date Survey Completed<br><br>06/19/2023 |
|---|---|---|
| **Name of Provider or Supplier**<br><br>Nichols Mohs And Skin Surgery | **Street Address, City, State**<br><br>303 N Keene St, Ste 201, Columbia, MO | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---|---|
| **D5217** | EVALUATION OF PROFICIENCY TESTING PERFORMANCE<br>CFR(s): 493.1236(c)(1)<br><br>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.<br><br>This STANDARD is not met as evidenced by:<br>Based on review of proficiency records for 2021/2022 and interview with the laboratory manager, the laboratory failed to establish a means to verify the accuracy of potassium hydroxide (KOH) microscopic examination testing twice a year.<br>Findings: 1. Review of proficiency records for 2021 and 2022 showed the laboratory failed to prove accuracy on the non-regulated analyte for KOH microscopic exams twice yearly. 2. Interview with the laboratory manager on June 19, 2023 at 11:30 AM confirmed the laboratory failed to verify the accuracy of the non-regulated KOH microscopic exam testing twice annually. |