

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  26D2069553	<b>(X3) Date Survey Completed</b>  01/11/2022
<b>Name of Provider or Supplier</b>  Missouri Dermatology Laser & Vein Center	<b>Street Address, City, State</b>  1011 Bowles, Suite 121, Fenton, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation of Surgipath tissue marking and margin dyes, potassium hydroxide (KOH) reagent and interview with testing personnel (TP) #3, the laboratory failed to discard expired tissue marking dyes and KOH reagent. Findings: 1. Observation of Surgipath tissue marking and margin dyes showed: Black tissue marking dye lot # 09132017, expiration date September 2018 still in use. Yellow tissue marking dye lot # 3801754, expiration date March 25, 2020 still in use. Green tissue marking dye lot # 032719, expiration date March 27, 2020 still in use. 2. Observation of Health Link potassium hydroxide (KOH) reagent showed lot # 132816, expiration date September 25, 2014 still in use. 3. Interview with TP #3 on January 11, 2022 at 2:00 PM confirmed the laboratory failed to discard expired tissue marking dyes and KOH reagent.</p>