

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 26D2074422	(X3) Date Survey Completed 08/14/2018
Name of Provider or Supplier Washington University Metabolomics	Street Address, City, State 425 S Euclid Ave, Rm 10511, Saint Louis, MO	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6128	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least annually after the first year, unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on review of personnel documentation and interview with the laboratory director, the technical supervisor failed to perform 2 of 2 competency evaluations for 2017. Findings: 1. Review of 2017 employee competencies revealed the technical supervisor failed to perform 2 of 2 competencies for testing personnel of high complexity testing. 2. Interview with the laboratory director on August 14, 2018 at 10:00PM confirmed the technical supervisor failed to perform annual competencies for 2017.</p>