

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 27D0689518	(X3) Date Survey Completed 04/26/2021
Name of Provider or Supplier Wheatland Memorial Healthcare Laboratory	Street Address, City, State 530 3rd Street Northwest, Harlowton, MT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5403	<p>PROCEDURE MANUAL CFR(s): 493.1251(b)</p> <p>The procedure manual must include the following when applicable to the test procedure: (1) Requirements for patient preparation; specimen collection, labeling, storage, preservation, transportation, processing, and referral; and criteria for specimen acceptability and rejection as described in 493.1242. (2) Microscopic examination, including the detection of inadequately prepared slides. (3) Step-by-step performance of the procedure, including test calculations and interpretation of results. (4) Preparation of slides, solutions, calibrators, controls, reagents, stains, and other materials used in testing. (5) Calibration and calibration verification procedures. (6) The reportable range for test results for the test system as established or verified in 493.1253. (7) Control procedures. (8) Corrective action to take when calibration or control results fail to meet the laboratory's criteria for acceptability. (9) Limitations in the test methodology, including interfering substances. (10) Reference intervals (normal values). (11) Imminently life-threatening test results, or panic or alert values. (12) Pertinent literature references. (13) The laboratory's system for entering results in the patient record and reporting patient results including, when appropriate, the protocol for reporting imminently life threatening results, or panic, or alert values. (14) Description of the course of action to take if a test system becomes inoperable.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory procedures, reports and interview with technical supervisor (TS) #1, the laboratory failed to have a step by step procedure for microscopic urinalysis; failed to include in their procedure manuals the reference intervals (normal values) for microscopic urinalysis; and failed to include the normal reference range in the patient results reports. Findings: 1. No step by step procedure for microscopic urinalysis was available for review 2. No reference intervals (normal values) were available in the laboratory procedure manual for microscopic urinalysis. 3. The laboratory patient results report lacked normal reference intervals for</p>

microscopic urinalysis. 4. Interview with the TS #1 on 4/26/2021 at 10:00 AM, confirmed the laboratory failed to have a step-by-step procedure of microscopic urinalysis and failed to include normal values for microscopic urinalysis in their reports and procedure manual.

D5439

CALIBRATION AND CALIBRATION VERIFICATION

CFR(s): 493.1255(b)

Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.

This STANDARD is not met as evidenced by:

Based on review of the calibration records for the Siemens Dimension EXL 200 chemistry analyzer for the analytes sodium, potassium, and chloride and interview with the technical supervisor (TS) #1, the laboratory failed to perform at least a three point (a minimal, mid-point, and maximum) calibration verification every six months. Findings: 1. Review of 2019 and 2020 calibration records for the Siemens Dimension EXL 200 analyzer for the analytes: sodium, potassium, and chloride, revealed the laboratory failed to perform a calibration including, at least, a minimal, midpoint, and maximum value for each analyte, every six months. 2. Interview with the TS #1 on 4/26/2021, 2021 at 10:45 AM confirmed the laboratory failed to perform at least a three-point calibration for sodium, potassium, and chloride on the Siemens Dimension EXL 200 analyzer every six months.