Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 27D0929493	(X3) Date Survey Completed 04/05/2021
Name of Provider or Supplier Bozeman Health Belgrade Clinic	**Street Address, City, State** 206 Alaska Frontage Road, Belgrade, MT	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports. This STANDARD is not met as evidenced by: Based on review of laboratory records and interview with general supervisor (GS) #1, the laboratory failed to record temperature for the dry heat block (Fisher Scientific Isotemp Dry Bath). Findings: 1. No temperature log for the dry heat block was available for review. 2. Interview with GS #1 on 4/5/2021 at 11:00 AM confirmed the laboratory failed to document temperatures for the dry heat block.