

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0453381	(X3) Date Survey Completed 11/07/2022
Name of Provider or Supplier Bellevue Health & Emergency Clinic Inc	Street Address, City, State 1002 W Mission Ave - 2nd Floor, Bellevue, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on off-site desk review of the Proficiency Testing (PT) data report (CASPER Report 155) and graded results from the proficiency testing organization American Proficiency Institute (API), the laboratory failed to successfully participate in proficiency testing for the analyte ABO/D(Rho). Refer to D2163.</p>
D2163	<p>ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(g)</p> <p>Failure to achieve an overall testing event score of satisfactory for two consecutive</p>

testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a review of the Proficiency Testing (PT) data report (CASPER Report 155) and graded results from the proficiency testing organization American Proficiency Institute (API), the laboratory failed to successfully participate in ABO/D(Rho), in two out of three testing events. Findings are: 1. API 2021 Event 3 for ABO/D(Rho) the score was 80%. 2. API 2022 Event 2 for ABO/D(Rho) the score was 0%.