

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0453431	(X3) Date Survey Completed 07/01/2021
Name of Provider or Supplier Memorial Community Hospital	Street Address, City, State 810 North 22nd Street, Blair, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5775	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of the laboratory's list of tests performed, lack of documentation, and interview with the general supervisor the laboratory failed to have a system that twice annually evaluated the relationship between test results using different methodologies for white blood cell (WBC) differentials for 2020. Findings are: 1. Review of the laboratories list of tests performed revealed the laboratory was performing WBC differentials by automated instrument and by manual method. 2. No documentation could be presented at the time of survey to demonstrate the laboratory evaluated the relationship between these different methodologies. 3. Interview with the general supervisor on 7/1/2021 at 1:35 PM confirmed patient testing was being performed using both methods, but no semi annual evaluations had been documented for 2020 comparing these two methodologies.</p>