

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0454924	(X3) Date Survey Completed 09/24/2018
Name of Provider or Supplier York General Health Care Services	Street Address, City, State 2222 North Lincoln Avenue, York, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: The laboratory failed to achieve satisfactory scores for the immunohematology analyte Compatibility Testing for the third event 2017 and the second events 2018 (see D2173). This results in the unsuccessful performance in proficiency testing for this analyte.</p>
D2173	<p>COMPATIBILITY TESTING CFR(s): 493.863(a)</p> <p>Failure to attain an overall testing event score of at least 100 percent is unsatisfactory</p>

performance.

This STANDARD is not met as evidenced by:

Based on desk review of proficiency testing for 2017 and 2018, this laboratory had unsatisfactory performance for the analyte Compatibility Testing for the third event 2017 (score 80%) and the second event 2018 (score 60%).