

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0455238	(X3) Date Survey Completed 08/25/2021
Name of Provider or Supplier Gynecologic Surgeons & Obstetricians Pc	Street Address, City, State 6050 Village Drive, Lincoln, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on desk review of proficiency testing the laboratory failed to achieve satisfactory scores for the subspecialty of bacteriology for the 2020 third event and 2021 second event. See D2028.
D2028	BACTERIOLOGY CFR(s): 493.823(e) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is

unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on proficiency testing (PT) desk review and the laboratory's graded PT results from American Proficiency Institute (API), the laboratory failed to achieve successful performance for the subspecialty bacteriology, in two out of three testing events.

Findings are: 1. 2020 third event subspecialty - bacteriology, score 0% 2. 2021 second event subspecialty - bacteriology, score 60%