

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0455307	(X3) Date Survey Completed 01/30/2018
Name of Provider or Supplier Physicians Laboratory Services Inc D/B/A	Street Address, City, State 7441 O Street Suite 100, Lincoln, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: The laboratory failed to achieve satisfactory scores for the general immunology analyte antinuclear antibodies (ANA) for the second and third events 2017 (see D2075). This results in the unsuccessful performance in proficiency testing for this analyte.
D2075	GENERAL IMMUNOLOGY CFR(s): 493.837(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte

in each testing event is unsatisfactory analyte performance for the testing event.

This STANDARD is not met as evidenced by:

Based on desk review of proficiency testing for 2017, this laboratory had unsatisfactory performance for the general immunology analyte ANA for the second event 2017 (score 40%) and the third event 2017 (score 40%).