

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0455760	(X3) Date Survey Completed 09/18/2020
Name of Provider or Supplier Grand Island Clinic Inc	Street Address, City, State 2444 W Faidley Ave, Grand Island, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on review of the laboratory's list of tests performed and interview with technical consultant, in 2019 the laboratory failed to enroll in an approved proficiency testing program for Human Chorionic Gonadotropin, qualitative serum and Rheumatoid Factor, qualitative serum. Findings are: 1. Review of the laboratory's list of tests performed revealed the laboratory performed two hundred and seventy five moderate complexity Human Chorionic Gonadotropin, qualitative serum and seventy five moderate complexity Rheumatoid Factor, qualitative serum in 2019. 2. Interview with technical consultant on 9/9/2020 at 3:00 PM confirmed the laboratory had not enrolled in 2019 proficiency testing for moderate complexity Human Chorionic Gonadotropin, qualitative serum and for moderate complexity Rheumatoid Factor, qualitative serum.</p>