

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0456376	(X3) Date Survey Completed 08/31/2021
Name of Provider or Supplier Perkins County Community Hospital	Street Address, City, State 900 Lincoln Avenue, Grant, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2000	<p>ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801</p> <p>Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994.</p> <p>This CONDITION is not met as evidenced by: Based on review of the laboratory's list of tests performed, review of 2020 proficiency testing, review of 2021 proficiency testing, and interview with general supervisor the laboratory failed to enroll in an approved proficiency testing program for moderate complexity Human Chorionic Gonadotropin (HCG), quantitative serum. Findings are: 1. Review of the laboratory's list of tests revealed the laboratory performed moderate complexity Human Chorionic Gonadotropin (HCG), quantitative serum testing. 2. Review of the laboratory's 2020 proficiency testing and 2021 proficiency testing revealed the laboratory failed to enroll in an approved program for moderate complexity Human Chorionic Gonadotropin (HCG), quantitative serum. 3. Interview with the general supervisor on 8/31/2021 at 3:03 PM confirmed the laboratory had not enrolled in an approved program for moderate complexity Human Chorionic Gonadotropin (HCG), quantitative serum for 2020 and 2021. 4. Interview with the general supervisor on 8/31/2021 at 3:03 PM confirmed the laboratory performed sixty-seven moderate complexity Human Chorionic Gonadotropin (HCG) quantitative serum tests from January 2020 to August 2021.</p>