

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0679425	(X3) Date Survey Completed 02/06/2023
Name of Provider or Supplier Beatrice Community Hospital & Health Center	Street Address, City, State 4800 Hospital Parkway, Beatrice, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on an off-site proficiency testing desk review of the laboratory's 2022 American Proficiency Institute proficiency testing (PT) records and a telephone interview with the laboratory manager on February 6, 2023, it was determined the laboratory failed to successfully participate in proficiency testing for the analyte cell identification/white blood cell differential. Refer to D2130.</p>
D2130	<p>HEMATOLOGY CFR(s): 493.851(f)</p>

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on desk review of proficiency testing for 2022, the laboratory's graded Proficiency Testing results from American Proficiency Institute, and a phone interview with the laboratory manager the laboratory had an unsatisfactory performance for the analyte Cell Identification/White Blood Cell Differential.

Findings are: 1. 2022 first event, analyte - Cell Identification/White Blood Cell Differential, score 20%. 2. 2022 third event, analyte - Cell Identification/White Blood Cell Differential, score 68%. 3. Phone interview with the laboratory manager on February 6, 2023 at 10:45 AM confirmed the laboratory had an unsatisfactory performance for the analyte Cell Identification/White Blood Cell Differential.