

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0684274	(X3) Date Survey Completed 01/09/2018
Name of Provider or Supplier Braddock Finnegan Helget Dermatology Pc	Street Address, City, State 7911 West Center Road, Omaha, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's list of tests performed, lack of documentation and staff interview at 9:40 AM on 1/9/2018, the laboratory failed to verify the accuracy for dermatopathology. Findings are: 1. Review of the laboratory's list of tests performed revealed the laboratory performing high complexity dermatopathology testing by three testing personnel. 2. No documentation could be presented through proficiency testing or laboratory comparison to verify the accuracy of this testing for 2016 and 2017. 3. Staff interview revealed the laboratory had not enrolled in proficiency testing or laboratory comparisons for the previous 2 years for the three individuals performing this testing.</p>