

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D0697020	(X3) Date Survey Completed 11/17/2025
Name of Provider or Supplier Gordon Memorial Hospital	Street Address, City, State 300 East 8th Street, Gordon, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review was completed on November 17, 2025. At the time of the review, the laboratory was not in compliance with the Clinical Laboratory Improvement Amendments of 1988, 42 CFR 493.1 through 42 CFR 493.1780. The following condition deficiencies were cited: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6076 - 42 C.F.R. 493.1441 Condition: Laboratories performing high complexity testing; laboratory director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a desk review of proficiency testing (PT) records from the Certification and Survey Provider Enhanced Reporting (CASPER) 0155 report and American</p>

	<p>Proficiency Institute (API) 2025 proficiency testing records, the laboratory did not successfully participate in a proficiency testing program approved by HHS, for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. The laboratory failed to successfully participate in the specialty of Routine Chemistry for the analyte Iron, Total. Refer to D2096.</p>
<p>D2096</p>	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p> <p>(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of CASPER 0155 report, American Proficiency Institute (API) proficiency testing 2025 event 2 and event 3 records, and an interview with the laboratory manager, the laboratory failed to achieve an overall satisfactory performance (80% or better) for the same analyte in two of three consecutive testing events in specialty of Routine Chemistry for the analyte Iron, Total. Findings are: 1. Review of the CASPER 0155 report revealed the following results: Iron, Total Event 2: The laboratory received an unsatisfactory score of 60%. Iron, Total Event 3: The laboratory received an unsatisfactory score of 40%. 2. A review of the API 2025 proficiency testing records confirmed the laboratory received the above results. 3. Phone interview with the laboratory manager on 11/17/2025 at 4:15 PM confirmed the laboratory received the above results.</p>
<p>D6076</p>	<p>LABORATORY DIRECTOR CFR(s): 493.1441</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on a proficiency testing desk review of CASPER 0155 report and American Proficiency Institute 2025 event 2 and event 3 records, the laboratory director failed to manage successful proficiency testing participation. Refer to D6089.</p>
<p>D6089</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1445(e)(4)(i)</p> <p>(e)(4)(i) The proficiency testing samples are tested as required under subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on a proficiency testing desk review of CASPER 0155 report and American Proficiency Institute 2025 event 2 and event 3 records, the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2096.</p>