

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D1008576	(X3) Date Survey Completed 05/03/2021
Name of Provider or Supplier Chi Health Nebraska Heart	Street Address, City, State 7500 South 91st Street, Lincoln, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on proficiency testing desk review, the laboratory failed to successfully participate in proficiency testing for the analyte partial oxygen (pO2) blood gas. See D2088.</p>
D2088	<p>ROUTINE CHEMISTRY CFR(s): 493.841(b)</p> <p>Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p>

This STANDARD is not met as evidenced by:

Based on proficiency testing (PT) desk review and the laboratory's graded PT results from College of American Pathologists (CAP), the laboratory failed to achieve successful performance for the analyte, partial oxygen (pO₂) blood gas, in two out of three testing events. Findings are: 1. 2020 second event, analyte - pO₂ blood gas, score 20% 2. 2021 first event, analyte - pO₂ blood gas, score 60%