

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D1041326	(X3) Date Survey Completed 04/23/2025
Name of Provider or Supplier Dermatology Specialists Of Omaha Llc	Street Address, City, State 909 N 96th Street Suite 201, Omaha, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's list of tests performed, lack of verification records, and interview with testing personnel #14, as indicated on the CMS-209, the laboratory failed to have a system for verifying the accuracy of histology biopsy testing at least twice annually for 2023 and 2024. Findings are: 1. Review of the laboratory's list of tests performed revealed the laboratory performs histology biopsy testing. 2. Interview with testing personnel #14, as indicated on the CMS-209, on 4/23 /2025 at 11:44 AM confirmed the laboratory did not perform accuracy verification for histology biopsy testing in 2023 and 2024.</p>