

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 28D1056933	(X3) Date Survey Completed 05/19/2026
Name of Provider or Supplier Children's Physicians - Lavista	Street Address, City, State 10705 Hillcrest Plaza, Lavista, NE	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5787	<p>TEST RECORDS CFR(s): 493.1283(a)</p> <p>(a) The laboratory must maintain an information or record system that includes the following: (a)(1) The positive identification of the specimen. (a)(2) The date and time of specimen receipt into the laboratory. (a)(3) The condition and disposition of specimens that do not meet the laboratory's criteria for specimen acceptability. (a)(4) The records and dates of all specimen testing, including the identity of the personnel who performed the test(s).</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of 2025 complete blood count (CBC) patient result strips and interview with the laboratory director the laboratory failed to maintain positive identification of specimen on four out of four patient result strips. Findings are: 1. Surveyor review of 2025 CBC patient result strips from 6/23/2025, 7/7/2025, and 7/16/2025 revealed three patient results strips had a different medical record number (MRN) than the patient sticker attached to the patient result strip. 2. Surveyor review of 2025 CBC patient result strip from 2/7/2025 revealed one patient result strip had no MRN on the CBC patient result strip. 3. Interview with the lab director on 5/19/2026 at 10:32 AM confirmed the findings above.</p>