

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  29D0881077	<b>(X3) Date Survey Completed</b>  06/29/2018
<b>Name of Provider or Supplier</b>  Aistheta Reno Medical Skin Care Center	<b>Street Address, City, State</b>  500 Damonte Ranch Pkwy Ste 1056, Reno, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	<p>This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on 6/29/18. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.</p>
<b>D5315</b>	<p><b>SPECIMEN SUBMISSION, HANDLING, AND REFERRAL</b> CFR(s): 493.1242(c)</p> <p>The laboratory must refer a specimen for testing only to a CLIA-certified laboratory or a laboratory meeting equivalent requirements as determined by CMS.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records and interview with the laboratory director, the laboratory did not have documentation that the reference laboratories used for the twice per year accuracy checks for the reading of histopathology slides were CLIA-certified laboratories or met equivalent requirements as determined by CMS. Findings include: Laboratory records reviewed did not indicate the CLIA numbers of the laboratories used for the twice per year accuracy checks for the reading of histopathology slides for Mohs and dermatopathology diagnosis. The laboratory director interviewed during the on-site survey on 6/29/18 at approximately 2:30 PM indicated that he would obtain the CLIA numbers of the laboratories but did not submit the information to the inspector to date. The laboratory performs approximately 400 histopathology tests annually.</p>