

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  29D0931553	<b>(X3) Date Survey Completed</b>  03/20/2019
<b>Name of Provider or Supplier</b>  Skin Cancer & Dermatology Institute-Gs Richards Bl	<b>Street Address, City, State</b>  3950 Gs Richards Blvd, Carson City, NV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	This Statement of Deficiencies was created as a result of an on-site CLIA recertification survey conducted at your facility on March 20, 2019. The findings and conclusions of any investigation by the Division of Public and Behavioral Health shall not be construed as prohibiting any criminal or civil investigations, actions or other claims for relief that may be available to any party under applicable federal, state, or local laws.
<b>D5203</b>	<p><b>SPECIMEN IDENTIFICATION AND INTEGRITY</b> CFR(s): 493.1232</p> <p>The laboratory must establish and follow written policies and procedures that ensure positive identification and optimum integrity of a patient's specimen from the time of collection or receipt of the specimen through completion of testing and reporting of results.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Mohs accession log, Mohs maps, slides and the electronic medical records (EMR), and interview with the laboratory manager, the laboratory failed to follow policies and procedures to ensure positive identification of a patient's specimen from collection to final report. Findings include: 1. A sample survey of 10 random Mohs cases from July 2017 to March 2019 revealed that for case, #MC17-727, the Mohs accession number in the EMR did not match the accession number in the accession log, Mohs map and slides. The accession number referenced in the EMR was MC17-733. 2. The laboratory manager interviewed during the on-site survey on 3/20/19 at approximately 2:30 PM confirmed the finding. The laboratory performs approximately 1,200 histopathology tests annually.</p>